

-----  
Midmark M11 - v1.0.0  
Total Cycles: 617

Name: MIDMARK  
Sterilizer ID: V2178198  
01/21/2021 10:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 111.3 F 0.0  
2:00 116.7 F 0.1  
4:00 155.0 F 0.2  
6:00 173.1 F 2.1  
8:00 195.3 F 6.0  
10:00 211.8 F 10.6  
12:00 222.8 F 5.7  
14:00 233.2 F 10.1  
16:00 234.1 F 8.2  
18:00 244.0 F 13.0  
20:00 241.8 F 11.5  
22:00 255.2 F 18.2  
24:00 266.8 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.0 F 28.5  
2:00 272.2 F 28.9  
3:00 272.5 F 29.2  
4:00 271.7 F 28.8

Min 271.5 F 28.4  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 175.3 F 0.0  
5:00 234.1 F 0.0  
10:00 232.6 F 0.0  
15:00 221.1 F 0.0  
20:00 217.3 F 0.0  
25:00 220.5 F 0.0  
30:00 229.7 F 0.0  
Drying Complete

FILLING: 1:00  
HEATING: 25:23  
STERILIZING: 4:00  
VENTING: 3:04  
DRYING: 30:00  
TOTAL CYCLE: 01:03:27  
CYCLE COMPLETE

01/21/2021 11:47 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_