
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000133
15:38 08/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:41
PRESSURIZING 3:42
STERILIZING 4:47
136.3 C 224kPa 4:47
136.4 C 225kPa 5:17
136.4 C 225kPa 5:47
136.4 C 226kPa 6:17
136.4 C 226kPa 6:47
136.4 C 223kPa 7:17
136.4 C 225kPa 7:47
136.4 C 224kPa 8:17
136.4 C 224kPa 8:17
VENTING 8:17
AIR DRYING 9:12
CYCLE COMPLETE 9:26

Digital Signature #
9EEBB676EE467B95
