
Midmark M11 - v1.0.0
Total Cycles: 2215

Name: RITTER 1
Sterilizer ID: V2178198
08/09/2022 11:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.2 F 0.0
2:00 200.5 F 4.4
4:00 204.3 F 9.0
6:00 225.4 F 8.1
8:00 230.0 F 8.7
10:00 240.5 F 11.8
12:00 238.5 F 10.9
14:00 237.2 F 8.8
16:00 250.1 F 15.5
18:00 261.5 F 21.9
20:00 272.1 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.3 F 29.3
2:00 272.0 F 28.8
3:00 271.2 F 28.4
4:00 271.9 F 29.2

Min 271.1 F 28.3
Max 272.4 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.5 F 0.0
5:00 231.6 F 0.0
10:00 228.1 F 0.0
15:00 224.3 F 0.0
20:00 226.7 F 0.0
25:00 228.5 F 0.0
30:00 238.8 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 20:37
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:16
CYCLE COMPLETE

08/09/2022 12:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____