

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1716

Name: RITTER 1  
Sterilizer ID: V2178198  
03/14/2022 03:07 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.6 F 0.0  
2:00 187.5 F 3.0  
4:00 189.4 F 7.3  
6:00 214.0 F 13.1  
8:00 226.5 F 7.8  
10:00 237.8 F 13.3  
12:00 239.0 F 10.6  
14:00 237.4 F 8.8  
16:00 250.6 F 15.7  
18:00 262.4 F 22.4  
20:00 272.7 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 272.4 F 29.1  
2:00 271.9 F 28.6  
3:00 271.2 F 28.4  
4:00 272.4 F 29.2

Min 271.1 F 28.2  
Max 272.5 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.2 F 0.0  
5:00 203.7 F 0.0  
10:00 220.1 F 0.0  
15:00 232.0 F 0.0  
20:00 229.5 F 0.0  
25:00 222.3 F 0.0  
30:00 218.3 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 20:21  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:58:31  
CYCLE COMPLETE

03/14/2022 04:05 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_