
Midmark M11 - v1.0.5
Total Cycles: 3453

Name: RITTER2
Sterilizer ID: V2313228
03/12/2024 03:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 115.5 F 0.0
2:00 118.2 F 0.3
4:00 137.9 F 2.4
6:00 168.5 F 6.1
8:00 188.1 F 10.7
10:00 213.3 F 4.4
12:00 225.0 F 9.2
14:00 234.0 F 10.5
16:00 239.0 F 11.3
18:00 236.9 F 9.6
20:00 250.6 F 16.1
22:00 262.2 F 22.9
24:00 272.7 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.2 F 29.3
2:00 271.5 F 29.4
3:00 271.9 F 29.2
4:00 271.9 F 29.5

Min 271.1 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.3 F 0.0
5:00 195.4 F 0.0
10:00 198.7 F 0.0
15:00 197.7 F 0.0
20:00 195.9 F 0.0
25:00 194.3 F 0.0
30:00 192.0 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 24:16
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:02:34
CYCLE COMPLETE

03/12/2024 04:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____