
Midmark M11 - v1.0.0
Total Cycles: 2861

Name: RITTER 1
Sterilizer ID: V2178198
02/06/2023 01:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.2 F 0.0
2:00 161.3 F 3.4
4:00 178.3 F 7.5
6:00 200.0 F 12.4
8:00 216.2 F 6.3
10:00 230.6 F 11.5
12:00 232.2 F 8.1
14:00 241.3 F 11.3
16:00 244.8 F 12.8
18:00 257.7 F 19.7
20:00 269.2 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.0 F 28.6
2:00 272.4 F 29.1
3:00 272.2 F 29.2
4:00 271.8 F 28.6

Min 271.6 F 28.5
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.6 F 0.0
5:00 233.8 F 0.0
10:00 229.8 F 0.0
15:00 230.8 F 0.0
20:00 226.8 F 0.0
25:00 227.9 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 20:58
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:15
CYCLE COMPLETE

02/06/2023 02:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____