
Midmark M11 - v1.0.0
Total Cycles: 4965

Name: RITTER 1
Sterilizer ID: V2178198
10/04/2024 09:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.7 F 0.0
2:00 152.7 F 0.5
4:00 159.4 F 1.5
6:00 180.2 F 4.8
8:00 201.9 F 9.4
10:00 222.1 F 9.9
12:00 223.7 F 7.7
14:00 235.2 F 12.6
16:00 236.4 F 9.5
18:00 238.4 F 9.4
20:00 247.3 F 13.8
22:00 259.2 F 20.1
24:00 269.4 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.2
1:00 272.0 F 28.8
2:00 272.2 F 28.6
3:00 271.3 F 28.2
4:00 272.4 F 29.0

Min 271.3 F 28.1
Max 272.4 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.7 F 0.0
5:00 206.7 F 0.0
10:00 194.4 F 0.0
15:00 193.0 F 0.0
20:00 224.0 F 0.0
25:00 239.7 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 24:56
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:03:24
CYCLE COMPLETE

10/04/2024 10:34 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____