

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001892  
11:31 25/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:49  
PRESSURIZING 1:57  
STERILIZING 2:52  
136.0 C 220kPa 2:52  
136.3 C 225kPa 3:22  
136.3 C 226kPa 3:52  
136.4 C 223kPa 4:22  
136.4 C 225kPa 4:52  
136.4 C 226kPa 5:22  
136.4 C 224kPa 5:52  
136.5 C 227kPa 6:22  
136.5 C 227kPa 6:22  
VENTING 6:22  
AIR DRYING 7:14  
CYCLE COMPLETE 31:57  
-----

Digital Signature #  
5FD49D87EA82A639  
-----