
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002147
9:56 16/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:33
PRESSURIZING 2:38
STERILIZING 3:39
136.0 C 220kPa 3:39
136.4 C 224kPa 4:09
136.2 C 223kPa 4:39
136.4 C 225kPa 5:09
136.4 C 223kPa 5:39
136.3 C 224kPa 6:09
136.4 C 224kPa 6:39
136.1 C 221kPa 7:09
136.1 C 221kPa 7:09
VENTING 7:10
AIR DRYING 8:01
CYCLE COMPLETE 38:01

Digital Signature #
CC1E82518E6DEE91
