
Midmark M11 - v1.0.5
Total Cycles: 3897

Name: RITTER2
Sterilizer ID: V2313228
08/27/2024 03:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.1 F 0.0
2:00 189.9 F 1.0
4:00 189.0 F 5.1
6:00 214.7 F 11.3
8:00 227.8 F 8.1
10:00 237.6 F 10.3
12:00 245.4 F 13.0
14:00 245.7 F 13.2
16:00 259.7 F 21.0
18:00 260.8 F 21.5
20:00 272.2 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 272.5 F 28.9
2:00 273.1 F 29.5
3:00 272.0 F 28.8
4:00 272.7 F 29.0

Min 271.7 F 28.5
Max 273.2 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 221.0 F 0.0
5:00 221.5 F 0.0
10:00 233.0 F 0.0
15:00 226.2 F 0.0
20:00 221.8 F 0.0
25:00 220.5 F 0.0
30:00 218.3 F 0.0
Drying Complete

FILLING: 2:25
HEATING: 20:08
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 00:59:32
CYCLE COMPLETE

08/27/2024 04:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____