
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005495
11:36 04/05/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:52
PRESSURIZING 1:55
STERILIZING 2:43
136.0 C 220kPa 2:43
136.4 C 224kPa 3:13
136.4 C 221kPa 3:43
136.4 C 223kPa 4:13
136.2 C 220kPa 4:43
136.4 C 225kPa 5:13
136.4 C 223kPa 5:43
136.4 C 223kPa 6:13
136.4 C 223kPa 6:13
VENTING 6:13
AIR DRYING 7:03
CYCLE COMPLETE 11:54

Digital Signature #
FEED12BCECE48ADC
