
Midmark M11 - v1.0.5
Total Cycles: 1928

Name: RITTER2
Sterilizer ID: V2313228
01/12/2023 12:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.1 F 0.0
2:00 183.2 F 3.6
4:00 187.2 F 8.3
6:00 208.8 F 10.3
8:00 222.8 F 8.8
10:00 235.1 F 9.9
12:00 241.7 F 12.9
14:00 242.6 F 12.2
16:00 256.2 F 19.4
18:00 268.3 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.6 F 29.9
2:00 271.4 F 29.2
3:00 272.4 F 30.1
4:00 272.1 F 29.6

Min 271.3 F 29.1
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.0 F 0.0
5:00 232.1 F 0.0
10:00 245.5 F 0.0
15:00 237.4 F 0.0
20:00 228.8 F 0.0
25:00 222.9 F 0.0
30:00 225.9 F 0.0

Drying Complete

FILLING: 1:22
HEATING: 19:11
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:57:41
CYCLE COMPLETE

01/12/2023 01:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____