
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002135
7:22 10/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:39
PRESSURIZING 2:45
STERILIZING 3:50
136.0 C 220kPa 3:50
136.4 C 224kPa 4:20
136.4 C 224kPa 4:50
136.4 C 224kPa 5:20
136.3 C 223kPa 5:50
136.2 C 222kPa 6:20
136.3 C 223kPa 6:50
136.4 C 223kPa 7:20
136.4 C 223kPa 7:20
VENTING 7:20
AIR DRYING 8:11
CYCLE COMPLETE 38:11

Digital Signature #
583A23C7533F8779
