

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 005105  
10:48 13/02/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:53  
PRESSURIZING 1:56  
STERILIZING 3:02  
136.4 C 220kPa 3:02  
136.4 C 220kPa 3:32  
136.5 C 225kPa 4:02  
136.2 C 218kPa 4:32  
136.4 C 222kPa 5:02  
136.4 C 224kPa 5:32  
136.4 C 219kPa 6:02  
136.4 C 222kPa 6:32  
136.4 C 222kPa 6:32  
VENTING 6:32  
AIR DRYING 7:23  
CYCLE COMPLETE 14:51  
-----

Digital Signature #  
B928B004AC3D59E7  
-----