

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3980

Name: RITTER2  
Sterilizer ID: V2313228  
09/30/2024 12:22 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 67.9 F 0.0  
2:00 70.8 F 0.0  
4:00 76.9 F 0.0  
6:00 103.9 F 0.1  
8:00 144.7 F 0.4  
10:00 172.6 F 1.7  
12:00 193.3 F 5.1  
14:00 207.7 F 8.6  
16:00 218.0 F 12.4  
18:00 223.7 F 6.2  
20:00 234.6 F 10.5  
22:00 234.6 F 8.5  
24:00 244.0 F 12.8  
26:00 242.1 F 11.4  
28:00 253.9 F 17.5  
30:00 264.3 F 23.7  
32:00 271.9 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.7 F 29.2  
2:00 271.6 F 28.7  
3:00 272.7 F 29.1  
4:00 271.6 F 28.5  
  
Min 271.5 F 28.4  
Max 272.8 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.2 F 0.0  
5:00 200.1 F 0.0  
10:00 199.9 F 0.0  
15:00 193.8 F 0.0  
20:00 190.0 F 0.0  
25:00 187.1 F 0.0  
30:00 185.1 F 0.0  
  
Drying Complete

FILLING: 0:52  
HEATING: 32:07  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 01:10:17  
CYCLE COMPLETE

09/30/2024 01:32 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_