

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000704  
9:48 28/10/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:40  
PRESSURIZING 2:46  
STERILIZING 3:54  
136.0 C 220kPa 3:54  
136.0 C 221kPa 4:24  
136.3 C 219kPa 4:54  
136.2 C 222kPa 5:24  
136.2 C 218kPa 5:54  
136.3 C 223kPa 6:24  
136.3 C 222kPa 6:54  
136.4 C 221kPa 7:24  
136.4 C 221kPa 7:24  
VENTING 7:24  
AIR DRYING 8:18  
CYCLE COMPLETE 8:27

-----  
Digital Signature #  
30CDD1FB6835A5E3  
-----