
Midmark M11 - v1.0.0
Total Cycles: 2846

Name: RITTER 1
Sterilizer ID: V2178198
02/02/2023 09:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.2 F 0.0
2:00 78.6 F 0.0
4:00 89.9 F 0.5
6:00 120.0 F 2.8
8:00 145.7 F 5.8
10:00 174.4 F 9.3
12:00 191.7 F 13.4
14:00 216.7 F 5.3
16:00 226.6 F 8.9
18:00 236.6 F 12.9
20:00 233.8 F 8.6
22:00 243.4 F 12.3
24:00 242.4 F 11.5
26:00 254.3 F 17.5
28:00 264.8 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.2 F 28.9
2:00 271.1 F 28.5
3:00 273.0 F 29.5
4:00 271.8 F 28.9

Min 271.0 F 28.2
Max 273.0 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.4 F 0.0
5:00 224.3 F 0.0
10:00 220.6 F 0.0
15:00 224.7 F 0.0
20:00 226.8 F 0.0
25:00 229.9 F 0.0
30:00 230.4 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 29:57
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:07:57
CYCLE COMPLETE

02/02/2023 10:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____