
Midmark M11 - v1.0.5
Total Cycles: 2284

Name: RITTER2
Sterilizer ID: V2313228
04/21/2023 01:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 139.4 F 0.0
2:00 141.2 F 1.8
4:00 161.6 F 4.8
6:00 188.3 F 9.0
8:00 206.2 F 13.5
10:00 219.8 F 6.8
12:00 231.0 F 11.3
14:00 231.8 F 8.2
16:00 242.7 F 12.3
18:00 244.2 F 12.8
20:00 256.2 F 19.1
22:00 266.9 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.9
1:00 272.2 F 29.7
2:00 272.0 F 29.1
3:00 271.7 F 29.4
4:00 272.3 F 29.4

Min 271.2 F 28.8
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.4 F 0.0
5:00 224.0 F 0.0
10:00 201.2 F 0.0
15:00 198.3 F 0.0
20:00 195.9 F 0.0
25:00 193.3 F 0.0
30:00 190.8 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 23:26
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:02:07
CYCLE COMPLETE

04/21/2023 02:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____