
Midmark M11 - v1.0.0
Total Cycles: 3602

Name: RITTER 1
Sterilizer ID: V2178198
08/01/2023 01:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.6 F 0.0
2:00 191.0 F 2.0
4:00 208.4 F 5.7
6:00 223.1 F 11.2
8:00 224.9 F 5.9
10:00 235.6 F 11.2
12:00 229.7 F 7.6
14:00 244.0 F 13.5
16:00 242.8 F 11.7
18:00 255.4 F 18.3
20:00 266.8 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.3 F 29.2
2:00 271.8 F 28.5
3:00 271.5 F 29.0
4:00 272.2 F 29.1

Min 271.0 F 28.3
Max 272.3 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 202.3 F 0.0
10:00 219.1 F 0.0
15:00 222.3 F 0.0
20:00 223.2 F 0.0
25:00 222.7 F 0.0
30:00 221.2 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 21:30
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:00:02
CYCLE COMPLETE

08/01/2023 02:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____