
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002039
13:41 13/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:02
PRESSURIZING 2:07
STERILIZING 3:07
136.4 C 225kPa 3:07
136.4 C 226kPa 3:37
136.4 C 226kPa 4:07
136.4 C 226kPa 4:37
136.2 C 224kPa 5:07
136.1 C 223kPa 5:37
136.4 C 226kPa 6:07
136.4 C 226kPa 6:37
136.4 C 226kPa 6:37
VENTING 6:37
AIR DRYING 7:28
CYCLE COMPLETE 37:28

Digital Signature #
400CBDBF8A3E404C
