
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000137
18:23 08/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:54
PRESSURIZING 3:54
STERILIZING 5:09
136.0 C 223kPa 5:09
136.4 C 226kPa 5:39
136.4 C 225kPa 6:09
136.3 C 224kPa 6:39
136.4 C 227kPa 7:09
136.3 C 225kPa 7:39
136.4 C 224kPa 8:09
136.4 C 226kPa 8:39
136.4 C 226kPa 8:39
VENTING 8:39
AIR DRYING 9:45
CYCLE COMPLETE 10:15

Digital Signature #
19CD705BFD554B81
