
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002036
11:18 13/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:51
PRESSURIZING 1:56
STERILIZING 3:02
136.5 C 226kPa 3:02
136.4 C 226kPa 3:32
136.4 C 225kPa 4:02
136.4 C 226kPa 4:32
136.4 C 227kPa 5:02
136.4 C 225kPa 5:32
136.2 C 223kPa 6:02
136.3 C 225kPa 6:32
136.3 C 225kPa 6:32
VENTING 6:32
AIR DRYING 7:22
CYCLE COMPLETE 33:26

Digital Signature #
D9C9DF13A3E4C362
