
Midmark M11 - v1.0.0
Total Cycles: 5419

Name: RITTER 1
Sterilizer ID: V2178198
03/17/2025 10:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 104.0 F 0.0
2:00 104.6 F 0.9
4:00 115.9 F 3.6
6:00 149.7 F 6.7
8:00 180.6 F 10.0
10:00 199.4 F 13.5
12:00 206.8 F 5.3
14:00 218.3 F 9.2
16:00 230.1 F 13.4
18:00 233.0 F 9.1
20:00 242.3 F 11.4
22:00 242.4 F 11.9
24:00 253.9 F 17.7
26:00 264.5 F 24.0
28:00 271.1 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.3
1:00 272.8 F 29.5
2:00 272.4 F 29.9
3:00 272.2 F 29.3
4:00 272.5 F 29.8

Min 271.8 F 29.2
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.9 F 0.0
5:00 189.6 F 0.0
10:00 196.1 F 0.0
15:00 193.6 F 0.0
20:00 200.5 F 0.0
25:00 208.4 F 0.0
30:00 211.0 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 28:53
STERILIZING: 4:00
VENTING: 3:53
DRYING: 30:00
TOTAL CYCLE: 01:07:34
CYCLE COMPLETE

03/17/2025 11:55 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____