
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001860
10:24 16/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:40
STERILIZING 3:40
136.0 C 220kPa 3:40
136.3 C 223kPa 4:10
136.1 C 221kPa 4:40
136.4 C 224kPa 5:10
136.4 C 225kPa 5:40
136.2 C 223kPa 6:10
136.4 C 225kPa 6:40
136.4 C 224kPa 7:10
136.4 C 224kPa 7:10
VENTING 7:10
AIR DRYING 8:04
CYCLE COMPLETE 8:23

Digital Signature #
C043681AACBCB5FC
