
Midmark M11 - v1.0.5
Total Cycles: 1345

Name: RITTER2
Sterilizer ID: V2313228
07/26/2022 09:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.2 F 0.0
2:00 154.2 F 0.1
4:00 171.6 F 0.7
6:00 209.4 F 5.2
8:00 220.5 F 10.3
10:00 230.2 F 7.0
12:00 237.5 F 11.8
14:00 236.6 F 9.7
16:00 238.8 F 9.1
18:00 249.3 F 15.5
20:00 261.5 F 22.3
22:00 272.4 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 272.0 F 29.5
2:00 272.4 F 29.4
3:00 271.4 F 28.9
4:00 272.4 F 29.8

Min 271.2 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.2 F 0.0
5:00 208.4 F 0.0
10:00 230.2 F 0.0
15:00 229.9 F 0.0
20:00 226.8 F 0.0
25:00 222.3 F 0.0
30:00 211.5 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 22:21
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:01:06
CYCLE COMPLETE

07/26/2022 10:09 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____