
Midmark M11 - v1.0.0
Total Cycles: 3978

Name: RITTER 1
Sterilizer ID: V2178198
11/03/2023 11:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.1 F 0.0
2:00 166.5 F 2.7
4:00 176.9 F 6.6
6:00 225.3 F 12.0
8:00 221.9 F 6.1
10:00 235.8 F 11.1
12:00 231.9 F 7.5
14:00 245.7 F 13.1
16:00 244.2 F 12.2
18:00 257.2 F 19.0
20:00 268.9 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.8 F 28.9
2:00 271.8 F 28.1
3:00 272.1 F 28.7
4:00 272.7 F 28.7

Min 271.2 F 28.0
Max 273.0 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.0 F 0.0
5:00 231.2 F 0.0
10:00 201.2 F 0.0
15:00 201.7 F 0.0
20:00 204.4 F 0.0
25:00 232.1 F 0.0
30:00 235.6 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 21:09
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 00:59:53
CYCLE COMPLETE

11/03/2023 12:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____