
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001939
14:17 09/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:42
PRESSURIZING 2:46
STERILIZING 3:49
136.5 C 224kPa 3:49
136.1 C 222kPa 4:19
136.2 C 222kPa 4:49
136.3 C 221kPa 5:19
136.0 C 222kPa 5:49
136.2 C 223kPa 6:19
136.2 C 223kPa 6:49
136.1 C 222kPa 7:19
136.1 C 222kPa 7:19
VENTING 7:19
CYCLE ABORTED 7:30
STOP BUTTON PRESSED
NOT STERILE
END TIME 7:30

Digital Signature #
023505060D907A6B
