
Midmark M11 - v1.0.5
Total Cycles: 4035

Name: RITTER2
Sterilizer ID: V2313228
10/16/2024 04:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.9 F 0.0
2:00 176.2 F 3.6
4:00 197.0 F 8.3
6:00 214.9 F 13.5
8:00 232.2 F 8.3
10:00 244.1 F 12.9
12:00 241.5 F 11.7
14:00 240.4 F 10.8
16:00 253.8 F 17.6
18:00 265.4 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.4 F 29.2
2:00 271.5 F 29.1
3:00 272.5 F 29.2
4:00 271.3 F 28.9

Min 271.2 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.3 F 0.0
5:00 234.8 F 0.0
10:00 215.1 F 0.0
15:00 196.4 F 0.0
20:00 179.9 F 0.0
25:00 176.5 F 0.0
30:00 174.3 F 0.0
Drying Complete

FILLING: 2:10
HEATING: 19:47
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 00:59:31
CYCLE COMPLETE

10/16/2024 05:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____