
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002107
10:59 02/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:10
PRESSURIZING 2:17
STERILIZING 3:23
136.0 C 221kPa 3:23
136.1 C 223kPa 3:53
136.2 C 225kPa 4:23
136.4 C 226kPa 4:53
136.2 C 224kPa 5:23
136.2 C 223kPa 5:53
136.2 C 223kPa 6:23
136.4 C 225kPa 6:53
136.4 C 225kPa 6:53
VENTING 6:53
AIR DRYING 7:44
CYCLE COMPLETE 37:44

Digital Signature #
59EE09C2066B2B96
