
Midmark M11 - v1.0.0
Total Cycles: 5473

Name: RITTER 1
Sterilizer ID: V2178198
04/03/2025 10:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.5 F 0.0
2:00 156.3 F 2.4
4:00 162.9 F 5.8
6:00 185.9 F 9.5
8:00 206.1 F 13.5
10:00 212.9 F 4.8
12:00 225.3 F 8.7
14:00 236.7 F 12.9
16:00 235.8 F 9.3
18:00 240.5 F 10.2
20:00 245.5 F 13.0
22:00 256.2 F 18.7
24:00 266.1 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.6 F 29.4
2:00 272.2 F 28.9
3:00 271.3 F 28.7
4:00 272.3 F 29.4

Min 271.3 F 28.5
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.3 F 0.0
5:00 202.5 F 0.0
10:00 206.6 F 0.0
15:00 202.8 F 0.0
20:00 206.7 F 0.0
25:00 244.1 F 0.0
30:00 223.2 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 25:47
STERILIZING: 4:00
VENTING: 4:00
DRYING: 30:00
TOTAL CYCLE: 01:04:39
CYCLE COMPLETE

04/03/2025 11:08 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____