
Midmark M11 - v1.0.5
Total Cycles: 1407

Name: RITTER2
Sterilizer ID: V2313228
08/24/2022 11:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.8 F 0.0
2:00 142.7 F 0.0
4:00 160.7 F 1.6
6:00 194.7 F 4.7
8:00 212.5 F 8.8
10:00 227.4 F 10.3
12:00 232.3 F 9.5
14:00 239.2 F 9.8
16:00 243.2 F 13.0
18:00 242.7 F 12.0
20:00 255.4 F 18.7
22:00 266.9 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.4 F 29.5
2:00 271.4 F 28.8
3:00 272.6 F 29.6
4:00 271.9 F 28.9

Min 271.4 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 207.7 F 0.0
10:00 221.2 F 0.0
15:00 229.0 F 0.0
20:00 229.2 F 0.0
25:00 229.4 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 23:30
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:01:29
CYCLE COMPLETE

08/24/2022 12:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____