
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000107
7:25 03/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:02
PRESSURIZING 4:02
STERILIZING 5:19
136.0 C 221kPa 5:19
136.3 C 222kPa 5:49
136.3 C 224kPa 6:19
136.4 C 225kPa 6:49
136.3 C 224kPa 7:19
136.3 C 225kPa 7:49
136.4 C 225kPa 8:19
136.3 C 225kPa 8:49
136.3 C 225kPa 8:49
VENTING 8:49
AIR DRYING 9:55
CYCLE COMPLETE 24:55

Digital Signature #
A6D4E49B5F389561
