

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1810

Name: RITTER2  
Sterilizer ID: V2313228  
12/06/2022 10:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 157.9 F 0.0  
2:00 156.7 F 0.1  
4:00 176.1 F 3.9  
6:00 196.5 F 9.2  
8:00 215.7 F 7.6  
10:00 228.8 F 10.2  
12:00 232.3 F 8.1  
14:00 241.9 F 11.5  
16:00 245.7 F 13.7  
18:00 258.6 F 20.9  
20:00 270.5 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.4  
1:00 272.6 F 29.7  
2:00 272.5 F 30.1  
3:00 272.3 F 29.5  
4:00 272.5 F 30.1  
  
Min 271.7 F 29.2  
Max 273.0 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.3 F 0.0  
5:00 204.6 F 0.0  
10:00 217.0 F 0.0  
15:00 223.5 F 0.0  
20:00 218.6 F 0.0  
25:00 244.7 F 0.0  
30:00 187.8 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 20:46  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:58:57  
CYCLE COMPLETE

12/06/2022 11:35 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02