
Midmark M11 - v1.0.5
Total Cycles: 3214

Name: RITTER2
Sterilizer ID: V2313228
12/14/2023 10:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.5 F 0.0
2:00 71.1 F 0.0
4:00 80.1 F 0.2
6:00 108.4 F 1.4
8:00 139.6 F 4.0
10:00 171.7 F 7.0
12:00 199.9 F 11.0
14:00 214.2 F 4.0
16:00 221.6 F 6.6
18:00 231.2 F 10.5
20:00 232.0 F 7.8
22:00 241.8 F 11.7
24:00 238.4 F 9.7
26:00 251.4 F 16.0
28:00 262.3 F 22.4
30:00 272.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.9
1:00 272.1 F 28.8
2:00 272.7 F 29.0
3:00 272.6 F 29.2
4:00 272.7 F 29.5

Min 271.7 F 28.8
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.0 F 0.0
5:00 232.3 F 0.0
10:00 222.1 F 0.0
15:00 223.2 F 0.0
20:00 212.9 F 0.0
25:00 217.5 F 0.0
30:00 219.4 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:02
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:09:06
CYCLE COMPLETE

12/14/2023 11:31 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____