
Midmark M11 - v1.0.5
Total Cycles: 424

Name: RITTER2
Sterilizer ID: V2313228
10/07/2021 06:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.6 F 0.0
2:00 73.8 F 0.0
4:00 77.3 F 0.0
6:00 98.6 F 0.1
8:00 129.8 F 0.2
10:00 164.4 F 2.4
12:00 204.8 F 6.0
14:00 204.5 F 9.5
16:00 217.7 F 13.2
18:00 230.5 F 7.8
20:00 240.2 F 12.0
22:00 237.2 F 9.3
24:00 240.0 F 10.1
26:00 247.4 F 13.9
28:00 258.5 F 20.0
30:00 268.5 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.5
1:00 272.4 F 29.0
2:00 271.9 F 29.2
3:00 271.3 F 28.5
4:00 272.0 F 28.7

Min 271.1 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.5 F 0.0
5:00 207.7 F 0.0
10:00 206.1 F 0.0
15:00 207.5 F 0.0
20:00 202.0 F 0.0
25:00 201.7 F 0.0
30:00 202.1 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 31:05
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:09:06
CYCLE COMPLETE

10/07/2021 07:59 AM
ACCEPTED: X
REJECTED:
Operator: U02