
Midmark M11 - v1.0.0
Total Cycles: 1581

Name: RITTER 1
Sterilizer ID: V2178198
02/01/2022 11:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.7 F 0.0
2:00 190.7 F 3.5
4:00 209.9 F 8.5
6:00 219.0 F 7.4
8:00 230.8 F 9.2
10:00 237.2 F 9.3
12:00 243.6 F 12.5
14:00 242.4 F 11.2
16:00 255.8 F 18.2
18:00 267.7 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.4 F 28.6
2:00 271.6 F 28.3
3:00 272.4 F 28.6
4:00 271.4 F 28.0

Min 271.4 F 28.0
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.3 F 0.0
5:00 233.9 F 0.0
10:00 226.2 F 0.0
15:00 231.0 F 0.0
20:00 225.5 F 0.0
25:00 232.6 F 0.0
30:00 231.2 F 0.0

Drying Complete

FILLING: 1:48
HEATING: 19:18
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:21
CYCLE COMPLETE

02/01/2022 12:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____