
Midmark M11 - v1.0.0
Total Cycles: 3397

Name: RITTER 1
Sterilizer ID: V2178198
06/15/2023 01:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.5 F 0.0
2:00 174.0 F 1.5
4:00 192.1 F 5.6
6:00 210.0 F 10.5
8:00 221.4 F 6.0
10:00 234.0 F 10.7
12:00 234.1 F 7.3
14:00 243.9 F 12.5
16:00 244.0 F 12.5
18:00 256.5 F 19.0
20:00 267.5 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.5
1:00 272.4 F 29.3
2:00 271.8 F 28.6
3:00 272.4 F 29.3
4:00 272.2 F 28.9

Min 271.3 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.5 F 0.0
5:00 212.0 F 0.0
10:00 221.8 F 0.0
15:00 223.9 F 0.0
20:00 223.7 F 0.0
25:00 222.4 F 0.0
30:00 221.5 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 21:15
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 00:59:31
CYCLE COMPLETE

06/15/2023 02:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____