
Midmark M11 - v1.0.0
Total Cycles: 3344

Name: RITTER 1
Sterilizer ID: V2178198
06/06/2023 01:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.3 F 0.0
2:00 188.9 F 0.3
4:00 202.0 F 5.3
6:00 219.9 F 10.7
8:00 225.6 F 6.4
10:00 239.1 F 12.4
12:00 238.9 F 10.6
14:00 237.5 F 9.4
16:00 252.4 F 16.6
18:00 264.5 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.5 F 28.9
2:00 271.6 F 28.5
3:00 272.5 F 29.0
4:00 271.7 F 28.3

Min 271.5 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.0 F 0.0
5:00 214.3 F 0.0
10:00 222.5 F 0.0
15:00 226.0 F 0.0
20:00 229.5 F 0.0
25:00 230.9 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 19:51
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:02
CYCLE COMPLETE

06/06/2023 02:20 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____