
Midmark M11 - v1.0.0
Total Cycles: 5503

Name:
Sterilizer ID: V2178198
04/14/2025 09:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.5 F 0.0
2:00 69.1 F 0.0
4:00 83.6 F 0.2
6:00 120.1 F 0.3
8:00 151.2 F 0.7
10:00 174.3 F 2.1
12:00 195.1 F 4.8
14:00 204.8 F 8.4
16:00 217.3 F 11.9
18:00 222.6 F 5.2
20:00 232.0 F 9.7
22:00 240.5 F 10.2
24:00 241.5 F 11.5
26:00 238.5 F 9.7
28:00 252.0 F 16.2
30:00 263.3 F 22.7
32:00 272.2 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.3 F 28.7
2:00 271.5 F 28.4
3:00 272.1 F 28.7
4:00 271.5 F 28.4

Min 271.4 F 28.1
Max 272.4 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.0 F 0.0
5:00 168.2 F 0.0
10:00 182.5 F 0.0
15:00 185.7 F 0.0
20:00 204.0 F 0.0
25:00 212.3 F 0.0
30:00 217.6 F 0.0

Drying Complete

FILLING: 1:01
HEATING: 32:10
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:10:51
CYCLE COMPLETE

04/14/2025 10:24 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____