
Midmark M11 - v1.0.5
Total Cycles: 2496

Name: RITTER2
Sterilizer ID: V2313228
06/09/2023 02:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.1 F 0.0
2:00 156.3 F 2.8
4:00 172.9 F 6.7
6:00 197.3 F 11.7
8:00 215.4 F 5.4
10:00 228.9 F 10.3
12:00 231.7 F 7.9
14:00 243.4 F 13.2
16:00 243.0 F 12.2
18:00 256.0 F 18.9
20:00 267.3 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.8 F 29.5
2:00 271.5 F 28.7
3:00 272.7 F 29.7
4:00 272.2 F 29.1

Min 271.4 F 28.6
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.0 F 0.0
5:00 205.4 F 0.0
10:00 212.4 F 0.0
15:00 215.3 F 0.0
20:00 216.5 F 0.0
25:00 224.5 F 0.0
30:00 226.0 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 21:25
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:54
CYCLE COMPLETE

06/09/2023 03:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____