
Midmark M11 - v1.0.5
Total Cycles: 2827

Name: RITTER2
Sterilizer ID: V2313228
09/05/2023 08:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.1 F 0.0
2:00 70.1 F 0.0
4:00 74.2 F 0.0
6:00 89.8 F 0.0
8:00 125.4 F 1.0
10:00 160.1 F 3.9
12:00 183.9 F 7.1
14:00 200.7 F 10.7
16:00 216.5 F 6.0
18:00 222.5 F 7.7
20:00 232.9 F 11.7
22:00 231.9 F 8.1
24:00 243.7 F 13.3
26:00 242.1 F 11.7
28:00 253.4 F 17.4
30:00 263.9 F 23.7
32:00 272.0 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.0 F 29.0
2:00 272.4 F 29.4
3:00 272.1 F 29.6
4:00 271.6 F 28.9

Min 271.5 F 28.8
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.8 F 0.0
5:00 213.7 F 0.0
10:00 220.3 F 0.0
15:00 221.0 F 0.0
20:00 227.0 F 0.0
25:00 228.9 F 0.0
30:00 231.6 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 32:07
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:10:11
CYCLE COMPLETE

09/05/2023 09:41 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____