
Midmark M11 - v1.0.5
Total Cycles: 2824

Name: RITTER2
Sterilizer ID: V2313228
08/31/2023 12:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.7 F 0.0
2:00 156.0 F 0.0
4:00 184.4 F 1.5
6:00 198.3 F 5.3
8:00 212.5 F 9.2
10:00 225.3 F 11.9
12:00 231.1 F 9.6
14:00 237.7 F 9.8
16:00 243.5 F 12.7
18:00 242.2 F 11.7
20:00 254.3 F 17.7
22:00 264.7 F 24.0
24:00 270.9 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.2
1:00 272.9 F 29.4
2:00 272.4 F 29.9
3:00 271.6 F 29.1
4:00 272.8 F 29.7

Min 271.5 F 29.0
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 172.7 F 0.0
5:00 235.5 F 0.0
10:00 203.2 F 0.0
15:00 182.8 F 0.0
20:00 177.4 F 0.0
25:00 174.5 F 0.0
30:00 172.4 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 24:41
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:03:08
CYCLE COMPLETE

08/31/2023 01:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____