
Midmark M11 - v1.0.0
Total Cycles: 1772

Name: RITTER 1
Sterilizer ID: V2178198
03/30/2022 03:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 208.2 F 0.0
2:00 192.9 F 3.9
4:00 222.3 F 9.1
6:00 227.1 F 7.9
8:00 238.1 F 10.7
10:00 233.3 F 8.0
12:00 241.0 F 10.6
14:00 246.7 F 13.6
16:00 259.9 F 20.6
18:00 271.2 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.3 F 28.6
2:00 271.8 F 28.7
3:00 272.0 F 28.4
4:00 271.8 F 28.7

Min 271.2 F 28.1
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.4 F 0.0
5:00 233.8 F 0.0
10:00 228.9 F 0.0
15:00 232.3 F 0.0
20:00 227.6 F 0.0
25:00 229.3 F 0.0
30:00 225.6 F 0.0
Drying Complete

FILLING: 2:10
HEATING: 18:39
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:58
CYCLE COMPLETE

03/30/2022 03:59 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____