

-----  
Midmark M11 - v1.0.5  
Total Cycles: 757

Name: RITTER2  
Sterilizer ID: V2313228  
02/03/2022 09:00 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 179.6 F 0.0  
2:00 176.0 F 2.7  
4:00 191.4 F 6.6  
6:00 210.8 F 11.9  
8:00 221.1 F 7.4  
10:00 234.3 F 13.1  
12:00 239.2 F 10.7  
14:00 237.0 F 9.3  
16:00 251.1 F 16.0  
18:00 262.4 F 22.6  
20:00 272.8 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 272.2 F 29.5  
1:00 272.6 F 29.8  
2:00 272.4 F 30.0  
3:00 272.3 F 29.5  
4:00 272.7 F 30.2  
  
Min 272.0 F 29.3  
Max 272.9 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.3 F 0.0  
5:00 227.7 F 0.0  
10:00 234.9 F 0.0  
15:00 226.1 F 0.0  
20:00 230.1 F 0.0  
25:00 225.6 F 0.0  
30:00 225.3 F 0.0  
Drying Complete

FILLING: 1:47  
HEATING: 21:09  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 00:59:60  
CYCLE COMPLETE

02/03/2022 09:59 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_