
Midmark M11 - v1.0.0
Total Cycles: 3575

Name: RITTER 1
Sterilizer ID: V2178198
07/25/2023 07:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.9 F 0.0
2:00 78.7 F 0.0
4:00 92.3 F 0.0
6:00 118.0 F 0.0
8:00 148.8 F 1.4
10:00 175.7 F 3.5
12:00 191.9 F 6.6
14:00 204.3 F 10.1
16:00 219.4 F 9.8
18:00 222.9 F 7.3
20:00 233.3 F 11.7
22:00 232.6 F 7.9
24:00 244.9 F 12.7
26:00 243.4 F 12.0
28:00 255.6 F 18.3
30:00 266.4 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.3 F 28.7
2:00 271.7 F 28.7
3:00 272.0 F 28.5
4:00 271.9 F 28.9

Min 271.1 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.2 F 0.0
5:00 203.0 F 0.0
10:00 201.7 F 0.0
15:00 201.4 F 0.0
20:00 198.7 F 0.0
25:00 197.5 F 0.0
30:00 198.7 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 31:38
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:09:33
CYCLE COMPLETE

07/25/2023 08:57 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____