

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3356

Name: RITTER 1  
Sterilizer ID: V2178198  
06/08/2023 09:03 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 145.7 F 0.0  
2:00 151.2 F 0.0  
4:00 172.5 F 0.4  
6:00 190.4 F 4.3  
8:00 206.8 F 8.6  
10:00 222.8 F 13.2  
12:00 226.9 F 7.7  
14:00 239.3 F 12.7  
16:00 236.7 F 9.8  
18:00 238.9 F 8.9  
20:00 247.8 F 14.3  
22:00 259.7 F 20.8  
24:00 270.2 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.6 F 29.2  
2:00 271.9 F 28.5  
3:00 271.9 F 29.0  
4:00 272.5 F 29.1

Min 271.5 F 28.3  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.3 F 0.0  
5:00 205.7 F 0.0  
10:00 220.6 F 0.0  
15:00 226.0 F 0.0  
20:00 230.1 F 0.0  
25:00 226.2 F 0.0  
30:00 222.9 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 24:56  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 01:03:02  
CYCLE COMPLETE

06/08/2023 10:05 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_