

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
1.0uS / 0.6ppm  
CYCLE NUMBER 001695  
15:03 07/06/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:05  
PRESSURIZING 3:06  
STERILIZING 4:04  
136.0 C 220kPa 4:04  
136.0 C 223kPa 4:34  
136.3 C 225kPa 5:04  
136.3 C 224kPa 5:34  
136.4 C 225kPa 6:04  
136.4 C 223kPa 6:34  
136.4 C 224kPa 7:04  
136.4 C 224kPa 7:34  
136.4 C 224kPa 7:34  
VENTING 7:34  
AIR DRYING 8:31  
CYCLE COMPLETE 8:33  
-----

Digital Signature #  
C96F04EE0DB5B94B  
-----