

-----  
Midmark M11 - v1.0.5  
Total Cycles: 764

Name: RITTER2  
Sterilizer ID: V2313228  
02/07/2022 01:31 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.3 F 0.0  
2:00 168.5 F 3.7  
4:00 186.5 F 8.3  
6:00 208.8 F 11.0  
8:00 223.8 F 8.7  
10:00 234.5 F 10.3  
12:00 241.1 F 12.5  
14:00 242.1 F 11.7  
16:00 255.2 F 18.6  
18:00 267.2 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.0  
1:00 272.3 F 29.4  
2:00 271.5 F 29.2  
3:00 272.4 F 29.4  
4:00 271.3 F 29.0  
  
Min 271.3 F 28.8  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.1 F 0.0  
5:00 233.1 F 0.0  
10:00 221.3 F 0.0  
15:00 220.1 F 0.0  
20:00 224.9 F 0.0  
25:00 231.6 F 0.0  
30:00 227.1 F 0.0  
  
Drying Complete

FILLING: 1:28  
HEATING: 19:23  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:57:60  
CYCLE COMPLETE

02/07/2022 02:28 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_