
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000586
13:16 16/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:46
PRESSURIZING 2:54
STERILIZING 4:03
136.0 C 229kPa 4:03
136.2 C 230kPa 4:33
136.2 C 230kPa 5:03
136.1 C 230kPa 5:33
136.3 C 231kPa 6:03
136.2 C 230kPa 6:33
136.4 C 233kPa 7:03
136.2 C 229kPa 7:33
136.2 C 229kPa 7:33
VENTING 7:33
AIR DRYING 8:26
CYCLE COMPLETE 23:26

Digital Signature #
B8036226D38BAF2C
