

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1379

Name: RITTER2  
Sterilizer ID: V2313228  
08/08/2022 05:03 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 153.5 F 0.0  
2:00 154.7 F 0.0  
4:00 183.6 F 0.7  
6:00 204.1 F 4.9  
8:00 217.8 F 9.7  
10:00 227.1 F 8.1  
12:00 237.2 F 11.3  
14:00 236.8 F 9.7  
16:00 237.8 F 9.3  
18:00 250.8 F 16.0  
20:00 262.9 F 23.0  
22:00 272.0 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 272.7 F 29.3  
2:00 271.5 F 28.7  
3:00 272.8 F 29.4  
4:00 271.8 F 28.7

Min 271.5 F 28.6  
Max 272.8 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.6 F 0.0  
5:00 205.0 F 0.0  
10:00 213.4 F 0.0  
15:00 209.9 F 0.0  
20:00 201.9 F 0.0  
25:00 202.8 F 0.0  
30:00 214.7 F 0.0  
Drying Complete

FILLING: 1:09  
HEATING: 22:08  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:00:34  
CYCLE COMPLETE

08/08/2022 06:03 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_